Results of ZERODUR spacer thickness ("8mm") measurement

Thickness and parallelism of the 18 pieces have been measured using a dial gage stand with 1μ graduation dial indicator as that been described in TN-18. Results are shown in table below.

Insert #	Thickness d – 8mm, (μ)	Nonparallelism (μ)		Insert #	Thickness d – 8mm, (μ)	Nonparallelism (μ)
81	−1.0	0.1		810	−1.0	0.2
82	−1.0	0.2		811	−1.5	0.2
83	−1.2	0.2		812	−1.0	0.1
84	−1.2	0.1		813	−1.0	0.1
85	−1.2	0.3		814	−1.1	0.1
86	−1.1	0.1		815	−1.1	0.1
87	−1.5	0.3		816	−0.9	0.1
88	−1.5	0.1		817	−1.0	0.1
89	−1.5	0.1		818	−1.0	0.2

Inserts ## 87, 88, 89, and 811 must be used together in a one plane.